sisco

Data export instructions

(applied to WINDOWS system)

Preparation:

- 1, Gas Detector 1 piece (Data storage is turned on and has history records)
- 2, Mini-USB cable 1 piece
- 3, One WINDOWS computer, equipped with WPS or Microsoft Forms (used to open data records).

A Detector preparations

The detector is powering on, enters the detection interface and then stands by.

B Computer preparations

The operating system of the host computer is XP or above (WIN10 operating system is recommended). Please contact supplier to get the data export software.

Install the software, as shown in the figure.



Follow the steps to complete the installation.

	1 Next > Cancel	С
Destination Folder	Browse	D
Space required:59.1MB Space available:19.2GB	can customize the installation path	
-	 Back Install Cancel	

Completing DataExport 1.1.1 Setup							
DataExport 1.1.1 has been installed on your computer. Click Finish to close Setup.							
Run DataExport 1.1.1 							

Step 3, complete the installation, and the icon 4 will appear on the desktop.



sisco

C Connect gas detector and computer by USB cable

1,Power on gas detector and the USB cable connect to computer

2, Double-click the desktop icon, such as the icon on the right, to open the software, and the following interface will appear.



D Export steps

①Select the serial port and click COM to preview the serial port.

*In this operation example, the computer shows that the serial port is COM4 (for the specific serial port number, please refer to the newly added corresponding detector hardware serial port in the device manager of the computer).

2 Select the baud rate, usually 38400.
4 Click "Scan Devices". 3 After setting, click "Open Seri al Port".

(5) Click the "File Location" option, and the "Save Concentration History Data" check box will pop up, select the location where the exported file is saved. *Double-click the gas item in the table below to customize the gas name (red box in the figure below), which can be exported to the document synchronously.

🍃 Datae	export tool					•		8
Options H	Help							
	Communication settings				Data expo	ort		
			Only supports multi-gas handheld detector protocol d					
Serial port:	COM4		Data	type	History data 🚺	Scanning	device 🤇	D
			Effec	tive channe	4			
Baud rate:	38400	2	Total	Total storage 827				
	[1	Expo	rt range	— 1 —	82	27 🛨	
Open serial port 3			Filel	File location D:\2022 08 01.xlsx				
						Read data		
		_		-				
	\frown	Cha	innel informa	ation				
Channel	Gas Density	Unit	Low alarm	High alarm	n Range	Decimal	ADC	
1	CO 0	PPM	50	100	1000	0	496	
-	0.0	1110	5.0	20.0	10.0		70.4	_
	Concentration hist	ory data preser	vation			×		
	← → ∨ ↑ ► >T	his PC>Local D	Disk©:)> v	Č Searc	h DataExport	Q		
	Organize 🔻 New folde	r			(EE 💌	0		
	* Quick access	Name	^	C	Date modified	^		
	Desktop	202108	06	2	2021-8-7 AM 11:2	5		
	Documents *	, .20		L		>		
	File name: 20	22_12_30						
	Save as type: (*.)	(IS,".XISX)						
	∧ Hide Folders			S	ave cance	el		

(Click "Start Export" to pop up a prompt box (as shown in the figure below) about the required export time, select "OK" and wait for the completion.

Prompt									
? It is expected to take 25 seconds to export data. Do you want to continue?									
Confirm	No								

E View export file

As shown in the figure, by default, the export operation date is used as the file name (year_month_day.xlsx). Office form software is required to open (WPS or Microsoft Excel).



